


PCT

WORLD INTELLECTUAL PROPERTY ORGANIZATION  
International Bureau

INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(51) International Patent Classification 6 : <b>G02B 21/00</b>	<b>A1</b>	(11) International Publication Number: <b>WO 99/44089</b> (43) International Publication Date: 2 September 1999 (02.09.99)
<p>(21) International Application Number: PCT/US99/04356</p> <p>(22) International Filing Date: 26 February 1999 (26.02.99)</p> <p>(30) Priority Data: 60/076,041 26 February 1998 (26.02.98) US</p> <p>(71) Applicant (for all designated States except US): THE GENERAL HOSPITAL CORPORATION [US/US]; 55 Fruit Street, Boston, MA 02114 (US).</p> <p>(72) Inventors; and (75) Inventors/Applicants (for US only): WEBB, Robert, H. [US/US]; 9 Old Concord Road, Lincoln, MA 01773 (US). TEARNEY, Guillermo, J. [US/US]; 118 Kinnaird St., Unit 3, Cambridge, MA 02139 (US). BOUMA, Brett, E. [US/US]; 12 Monmouth St., Quincy, MA 02171 (US).</p> <p>(74) Agent: LUKACHER, Kenneth, J.; Harris Beach &amp; Wilcox, LLP, 130 East Main Street, Rochester, NY 14604-1687 (US).</p>	<p>(81) Designated States: AL, AM, AT, AU, AZ, BA, BB, BG, BR, BY, CA, CH, CN, CU, CZ, DE, DK, EE, ES, FI, GB, GE, GH, GM, HR, HU, ID, IL, IN, IS, JP, KE, KG, KP, KR, KZ, LC, LK, LR, LS, LT, LU, LV, MD, MG, MK, MN, MW, MX, NO, NZ, PL, PT, RO, RU, SD, SE, SG, SI, SK, SL, TJ, TM, TR, TT, UA, UG, US, UZ, VN, YU, ZW, ARIPO patent (GH, GM, KE, LS, MW, SD, SL, SZ, UG, ZW), Eurasian patent (AM, AZ, BY, KG, KZ, MD, RU, TJ, TM), European patent (AT, BE, CH, CY, DE, DK, ES, FI, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, GW, ML, MR, NE, SN, TD, TG).</p> <p><b>Published</b> With international search report. Before the expiration of the time limit for amending the claims and to be republished in the event of the receipt of amendments.</p>	

(54) Title: CONFOCAL MICROSCOPY WITH MULTI-SPECTRAL ENCODING



(57) Abstract

A scanning confocal microscopy system, especially useful for endoscopy with a flexible probe which is connected to the end of an optical fiber (9). The probe has a grating (12) and a lens (14) which delivers a beam of multi-spectral light having spectral components which extend in one dimension across a region of an object and which is moved to scan in another dimension. The reflected confocal spectrum is measured to provide an image of the region.

**FOR THE PURPOSES OF INFORMATION ONLY**

Codes used to identify States party to the PCT on the front pages of pamphlets publishing international applications under the PCT.

AL	Albania	ES	Spain	LS	Lesotho	SI	Slovenia
AM	Armenia	FI	Finland	LT	Lithuania	SK	Slovakia
AT	Austria	FR	France	LU	Luxembourg	SN	Senegal
AU	Australia	GA	Gabon	LV	Latvia	SZ	Swaziland
AZ	Azerbaijan	GB	United Kingdom	MC	Monaco	TD	Chad
BA	Bosnia and Herzegovina	GE	Georgia	MD	Republic of Moldova	TG	Togo
BB	Barbados	GH	Ghana	MG	Madagascar	TJ	Tajikistan
BE	Belgium	GN	Guinea	MK	The former Yugoslav Republic of Macedonia	TM	Turkmenistan
BF	Burkina Faso	GR	Greece	ML	Mali	TR	Turkey
BG	Bulgaria	HU	Hungary	MN	Mongolia	TT	Trinidad and Tobago
BJ	Benin	IE	Ireland	MR	Mauritania	UA	Ukraine
BR	Brazil	IL	Israel	MW	Malawi	UG	Uganda
BY	Belarus	IS	Iceland	MX	Mexico	US	United States of America
CA	Canada	IT	Italy	NE	Niger	UZ	Uzbekistan
CF	Central African Republic	JP	Japan	NL	Netherlands	VN	Viet Nam
CG	Congo	KE	Kenya	NO	Norway	YU	Yugoslavia
CH	Switzerland	KG	Kyrgyzstan	NZ	New Zealand	ZW	Zimbabwe
CI	Côte d'Ivoire	KP	Democratic People's Republic of Korea	PL	Poland		
CM	Cameroon	KR	Republic of Korea	PT	Portugal		
CN	China	KZ	Kazakhstan	RO	Romania		
CU	Cuba	LC	Saint Lucia	RU	Russian Federation		
CZ	Czech Republic	LJ	Liechtenstein	SD	Sudan		
DE	Germany	LK	Sri Lanka	SE	Sweden		
DK	Denmark	LR	Liberia	SG	Singapore		
EE	Estonia						

## 5                    CONFOCAL MICROSCOPY WITH MULTI-SPECTRAL ENCODING

This application claims the priority benefit of U.S. Provisional Application No. 60/076,041, filed 26 February 1998.

Description

10            The present invention relates to systems (method and apparatus) for confocal microscopy for the examination or imaging of sections of a specimen of biological tissue, and particularly to such systems using multi-spectral illumination and processing of multi-spectral light.

          Currently, the use of fast scanning confocal microscopy is limited to accessible  
15            surfaces of the skin and the eye. The reason for this is that the only reliable methods for optical scanning must be performed in free space. In addition, the size of these optical scanners prohibit their use in small probes such as endoscopes or catheters. It is a feature of the invention to miniaturize the fast scanning mechanism and increase the number of medical applications of confocal microscopy to include all surfaces of the body, gynecologic  
20            applications, probe-based applications, and internal organ systems.

          Multi-spectral light was proposed for use in confocal microscopy, but only for imaging vertically-spaced regions of a body under examination. See B. Picard, U.S. Patent No. 4,965,441, issued October 25, 1990. An interferometer using a grating to obtain multi-spectral light which is resolved in the interferometer to obtain a spectroscopic image is  
25            disclosed in A. Knuttal, U.S. Patent 5,565,986, issued October 15, 1996. A lens having a color separation grating which obtains a multi-spectral light is disclosed in U.S. Patent No. 5,600,486, issued February 4, 1997. Such multi-spectral proposals are not effective for high resolution imaging using a compact, flexible probe. A confocal microscope system according to this invention can be miniaturized and incorporated into a compact probe. In addition, by  
30            allowing light delivery through a single optical fiber, the probe may also be easily incorporated into catheters or endoscopes. Thus, a confocal microscope in accordance with

-2-

5 the invention allows imaging of all accessible surfaces of the body and increases the biomedical applications of confocal microscopy by an order of magnitude.

Briefly described, a confocal microscopy system embodying the invention illuminates a region of interest in a body into which said probe may be inserted with a confocal spectrum extending along one dimension. Optics in said probe or physical movement of said probe enabled by attachment thereto of a flexible light conductive member (which may be an optical  
10 fiber), enables scanning of said spectrum along one or two additional dimensions thereby providing for two or three dimensional imaging of the region. The reflected confocal spectrum may be detected or decoded spectroscopically, preferably with a heterodyne detection mechanism which may be implemented interferometrically.

15 The invention will be more apparent from the following drawings wherein

Fig. 1 is a schematic diagram of a spectrally encoded confocal probe in accordance with the invention where specific wavelengths are shown for illustrative purposes, their exact values depending on the optical parameters of the system.

Fig. 2 is a plot of spectrally encoded light obtained by confocal detection using direct  
20 spectral detection in accordance with this invention, where different wavelengths are detected by turning the spectrometer grating.

Fig. 3 is a schematic diagram showing a system embodying the invention using a spectrometer for measurement of the spectrum,  $I(\lambda)$ , which corresponds to reflectance from different transverse locations,  $x$ , on the specimen.

25 Fig. 4 is a schematic diagram of a system embodying the invention having spectrally encoded confocal detection using interference spectroscopy.

Fig. 5A-D are schematic diagrams showing: (a) image formation; (b) translation of the optical fiber in the  $y$  direction; (c) rotation of the optical fiber in the forward firing mode; and (d) rotation of the optical fiber in the side firing mode.

-3-

5 Fig. 6 is a schematic diagram showing cross-sectional image formation by scanning the optical fiber or the objective lens along the z axis using a system embodying the invention.

Fig. 7 is another schematic diagram of a system embodying the invention wherein optical zoom is achieved by moving the focus of an intermediate lens in and out of the image plan of the objective.

10 Referring now to the figures, multi-spectral encoding for confocal microscopy uses a broad bandwidth source 10 as the input to the microscope. In the probe 8 of the microscope, the source spectrum provided via an optical fiber 9 is dispersed by a grating 12 and focused by an objective lens 14 onto the sample 16. A lens 9a is preferably disposed between the optical fiber 9 and the grating 12 to collimate the light from the optical fiber, as shown in Fig. 1, however, lens 9a may be removed. The spot for each wavelength is focused at a separate position, x, on the sample (Fig.1). The reflectance as a function of transverse location is determined by measuring the reflected confocal spectrum from the sample 16 returned from probe 8.

The number of wavelengths or points that may be resolved is determined by:

$$\frac{\lambda}{\delta\lambda} = mN, \quad (1)$$

20 where  $\lambda$  is the center wavelength,  $\delta\lambda$  is the bandwidth of the spectrum,  $N$  is the number of lines in the grating 12 illuminated by the polychromatic input beam 10, and  $m$  is the diffraction order. If the total bandwidth of the source is  $\Delta\lambda$ , the number of resolvable points,  $n$  is defined by:

$$n = \frac{\Delta\lambda}{\delta\lambda} \quad (2)$$

-4-

5 For an input source with a center wavelength of 800 nm, a bandwidth of 25 nm, an input spot diameter of 5 mm, a diffraction grating of 1800 lines/mm and a diffraction order of 1,  $n = 281$  points may be resolved by the spectrally encoded confocal system (FIG. 2). The parameters used in this example may be found in common, inexpensive optical components. The number of points may be increased by simply increasing the input spot diameter or the  
10 bandwidth of the source. Increasing the spot diameter increases the resultant probe diameter. Increasing the bandwidth of the source could be accomplished by using a broader bandwidth superluminescent diode, a rare earth doped fiber superfluorescent source, or a solid state modelocked laser.

Consider next the multi-spectral process. First, consider direct spectral measurement.  
15 The reflectance from the sample 16 as a function of transverse location is determined by measuring the reflected confocal spectrum from the sample arm 18. The spectrum may be measured efficiently by incorporating the probe 8 in the sample arm of a Michelson interferometer 20 (Fig. 3) and detecting the light transmitted through a high resolution spectrometer 21 at the output port 19 of the interferometer. Thus, each wavelength measured  
20 corresponds to a separate position,  $x$ , on the sample (Fig. 3). The advantage to this method over traditional real time confocal microscopy is that the fast axis scanning ( $\sim 15$  kHz) may be performed external to the probe 8 by the spectrometer 21 with approximately .1 nm spectral resolution for the parameters given above, well within reach of high quality spectrometers.

High sensitivity may be achieved through the use of heterodyne detection. If the  
25 reference arm 22 is modulated, such as by modulator 23 with mirror 24 (Fig. 3), the interference of light from the sample arm 18 and the reference arm 22 will also be modulated. High signal-to-noise ratios may be then achieved by lock-in detection on the reference arm modulation frequency of detector 26.

Another method for measuring the spectrum is interference or Fourier transform  
30 spectroscopy. This may be accomplished by inserting a linearly translating mirror 28 in the

-5-

5 reference arm 22 and measuring the cross-correlation output 30 from the interference spectrometer due to the interference of the reflected light from the sample and reference arms 18 and 22, respectively (Fig. 4). The advantages to this type of spectroscopic detection include the ability to achieve higher spectral resolutions than direct detection methods, efficient use of the returned light, inherent modulation of the reference arm 22 by the Doppler  
10 shift of the moving mirror 28, and the capability to extract both reflectance and phase data from the sample 16. The ability to extract phase data from the sample may allow detection of refractive index as a function of transverse position,  $x$ , which is useful to reveal the molecular composition of the sample as well as provide an additional source of image contrast other than the reflectivity of the sample specimen 16. Finally, interferometric detection has the potential  
15 to allow elimination of high order multiple scattering from the confocal signal by coherence gating.

Consider finally image formation. The multi-spectral encoding of the transverse location,  $x$ , allows the performance of a one-dimensional raster scan. To obtain an image, a scan of another axis must be performed, which is usually slower. Methods of accomplishing  
20 this slow scanning of the  $y$  axis include moving the optical fiber 9 in the  $y$  direction (Fig. 5B), or rotating the entire probe 8 around the optical fiber axis either in a forward scanning configuration (Fig. 5C) or a side-firing configuration (Fig. 5D). Cross-sectional images may be created by scanning the optical fiber 9 or the objective lens 14 along the  $z$  axis (Fig. 6). Finally, a zoom mode may be created by scanning the optical fiber 9 (or another lens 32  
25 between grating 12 and objective lens 14), in and out of the image plane of the objective lens (Fig. 7). Both linear motion along the  $y$  or  $z$  axis and rotation are easily accomplished in a compact probe by use of piezoelectric transducers. As shown in FIG. 5A, signals may be received by a computer 34 from spectroscopic detector 32 by a spectrometer (such as described in connection with FIG. 3) or Fourier transform (such as described connection with

-6-

- 5      FIG. 4) representing an image of the a microscopic section of the sample, and the image displayed on a display coupled to the computer.

From the foregoing description, it will be apparent that the invention provides a *confocal microscopy system which (a) is compact, optical fiber-based, capable of enabling confocal microscopy through a flexible catheter or endoscope; (b) is fast-scanning which takes*  
10      *place external to the probe; (c) allows phase information to be retrieved; and (d) provides a number of resolvable points proportional to the bandwidth of the source and the beam diameter on the grating. Variations and modifications in the herein described confocal microscopy system in accordance with the invention will undoubtedly suggest themselves to those skilled in the art. Accordingly, the foregoing description should be taken as illustrative*  
15      *and not in a limiting sense.*



-7-

5

Claims

1. A confocal microscope system which comprises a probe movable into a body region of interest, said probe having means for illuminating said region with a confocal spectrum of light extending along one dimension, means for obtaining an image of the region of the specimen by moving said spectrum along another dimension and measuring the reflected confocal spectrum of said light.  
10
2. The system according to Claim 1 wherein said probe is mounted on the end of a flexible, light-conducting member.
3. The system according to Claim 2 wherein said member is an optical fiber.
4. The system according to Claim 3 wherein said fiber is rotatable or translatable  
15 to move said probe in said another dimension.
5. The system according to Claim 1 wherein said means for moving said spectrum comprises means for moving an image plane containing said spectrum optically or by physically moving said probe.
6. The system according to Claim 5 wherein said probe is moved physically to  
20 scan said spectrum in said another dimension and said probe has means for optically moving said image plane to scan in still another direction, thereby enabling 3-D imaging.
7. The system according to Claim 1 wherein said means for obtaining said image comprises heterodyne detection means.
8. The system according to Claim 7 wherein said heterodyne detection means  
25 includes an interferometer.
9. The system according to Claim 8 wherein said interferometer has a sample arm terminated by said probe, a reference arm terminated by a mirror, an output arm having a spectroscopic detector, an input arm having a source of polychromatic illumination, and a beam splitter for directing light from said source to said sample and reference arms and  
30 directing interfering light containing said reflected confocal spectrum into said output arm.



-8-

- 5           10.    The system according to Claim 9 wherein said reference arm includes means for modulating said reflected spectrum.
11.    The system according to Claim 10 wherein said modulating means comprising means for reciprocally oscillating said mirror or a modulator.
12.    The system according to Claim 11 wherein said modulator or reciprocal  
10   oscillation is at a certain frequency, and means for lock-in operation of said detector at said frequency.
13.    The system according to Claim 9 wherein said detector is a spectrometer.
14.    The system according to Claim 9 wherein said detector includes a cross-correlator or a Fourier transform spectrometer.
- 15          15.    The system according to Claim 1 wherein said probe comprises a grating and an objective which provides said confocal spectrum in an image plan of said objective.
16.    The system according to Claim 15 wherein said probe is sufficiently small size to be insertable into an organ internal of said body.
17.    A system for confocally imaging tissue comprising:  
20               a source for producing light;  
              means for producing a confocal spectrum of said light;  
              means for focusing said confocal spectrum into said tissue, in which said confocal spectrum extends across said tissue at a focal plane, and receiving returned light from said tissue; and
- 25               means for detecting said returned light in accordance with spectrum of said returned light to provide an image representing said tissue.
18.    The system according to Claim 17 further comprising means for scanning said confocal spectrum in at least one dimension with respect to said tissue.
19.    The system according to Claim 17 wherein at least said producing means and  
30   said focusing and receiving means are located is a probe capable of insertion in a body.


-9-

- 5           20.    The system according to Claim 17 further comprising an optical fiber which provides said light from said source to said producing means, and provides said returned light from said focusing and receiving means to said detecting means.


1/7

FIG.1FIG.2

2/7




3/7



**FIG. 4**

4/7




5/7

ROTATING  
GRATING



**FIG.5C**




GRATING



**FIG.5B**

OPTICAL  
FIBER




Y AXIS  
TRANSLATION

ROTATING  
MIRROR AND  
GRATING



MIRROR



GRATING




OPTICAL  
FIBER




**FIG.5D**



6/7



**FIG.6**



**FIG. 7**

# INTERNATIONAL SEARCH REPORT

International application No.  
PCT/US99/04356

<b>A. CLASSIFICATION OF SUBJECT MATTER</b> IPC(6) : G02B 21/00 US CL : 359/368 According to International Patent Classification (IPC) or to both national classification and IPC														
<b>B. FIELDS SEARCHED</b> Minimum documentation searched (classification system followed by classification symbols) U.S. : 359/368,389 Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched Electronic data base consulted during the international search (name of data base and, where practicable, search terms used) APS <i>search terms: confocal, probe, microscope or endoscope, grating, diffraction</i>														
<b>C. DOCUMENTS CONSIDERED TO BE RELEVANT</b>														
Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.												
X	US 5,450,203 A (PENKETHMAN) 12 September 1995 (12/09/95), see entire document.	1-20												
<input type="checkbox"/> Further documents are listed in the continuation of Box C. <input type="checkbox"/> See patent family annex.														
<table border="0"> <tr> <td>* Special categories of cited documents:</td> <td>*T* later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention</td> </tr> <tr> <td>*A* document defining the general state of the art which is not considered to be of particular relevance</td> <td>*X* document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone</td> </tr> <tr> <td>*B* earlier document published on or after the international filing date</td> <td>*Y* document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art</td> </tr> <tr> <td>*L* document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)</td> <td>*G* document member of the same patent family</td> </tr> <tr> <td>*O* document referring to an oral disclosure, use, exhibition or other means</td> <td></td> </tr> <tr> <td>*P* document published prior to the international filing date but later than the priority date claimed</td> <td></td> </tr> </table>			* Special categories of cited documents:	*T* later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention	*A* document defining the general state of the art which is not considered to be of particular relevance	*X* document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone	*B* earlier document published on or after the international filing date	*Y* document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art	*L* document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)	*G* document member of the same patent family	*O* document referring to an oral disclosure, use, exhibition or other means		*P* document published prior to the international filing date but later than the priority date claimed	
* Special categories of cited documents:	*T* later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention													
*A* document defining the general state of the art which is not considered to be of particular relevance	*X* document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone													
*B* earlier document published on or after the international filing date	*Y* document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art													
*L* document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)	*G* document member of the same patent family													
*O* document referring to an oral disclosure, use, exhibition or other means														
*P* document published prior to the international filing date but later than the priority date claimed														
Date of the actual completion of the international search 24 JUNE 1999		Date of mailing of the international search report 20 JUL 1999												
Name and mailing address of the ISA/US Commissioner of Patents and Trademarks Box PCT Washington, D.C. 20231 Facsimile No. (703) 305-3230		Authorized officer MARK ROBINSON <i>[Signature]</i> Telephone No. (703) 305-3506												